
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001648
16:38 05/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:12
PRESSURIZING 2:15
STERILIZING 3:17
136.4 C 225kPa 3:17
136.3 C 225kPa 3:47
136.2 C 224kPa 4:17
136.2 C 224kPa 4:47
136.2 C 226kPa 5:17
136.4 C 226kPa 5:47
135.8 C 222kPa 6:17
136.4 C 225kPa 6:47
136.4 C 225kPa 6:47
VENTING 6:47
AIR DRYING 7:40
CYCLE COMPLETE 7:42

Digital Signature #
7A1139AE2DA86553
