

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000500  
11:10 07/05/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:02  
PRESSURIZING 3:06  
STERILIZING 4:21  
136.2 C 230kPa 4:21  
136.3 C 230kPa 4:51  
136.4 C 233kPa 5:21  
136.1 C 229kPa 5:51  
136.4 C 233kPa 6:21  
136.2 C 231kPa 6:51  
136.3 C 231kPa 7:21  
136.2 C 231kPa 7:51  
136.2 C 231kPa 7:51  
VENTING 7:51  
AIR DRYING 8:46  
CYCLE COMPLETE 23:46  
-----

Digital Signature #  
DCB4EFD316819798  
-----