
Midmark M11 - v1.0.5
Total Cycles: 1985

Name: RITTER2
Sterilizer ID: V2313228
01/31/2023 11:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.6 F 0.0
2:00 187.4 F 3.5
4:00 189.6 F 8.3
6:00 208.3 F 9.9
8:00 224.5 F 9.0
10:00 234.5 F 9.9
12:00 242.5 F 12.7
14:00 242.9 F 11.9
16:00 256.4 F 19.1
18:00 268.4 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.1 F 29.2
2:00 271.1 F 28.7
3:00 272.3 F 29.3
4:00 271.4 F 28.6

Min 271.1 F 28.5
Max 272.4 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.6 F 0.0
5:00 219.0 F 0.0
10:00 233.9 F 0.0
15:00 227.0 F 0.0
20:00 226.0 F 0.0
25:00 225.6 F 0.0
30:00 224.6 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 19:09
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:51
CYCLE COMPLETE

01/31/2023 12:06 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____