
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001699
11:39 07/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:38
STERILIZING 3:43
136.4 C 224kPa 3:43
136.1 C 223kPa 4:13
136.3 C 225kPa 4:43
136.1 C 222kPa 5:13
136.4 C 224kPa 5:43
136.3 C 225kPa 6:13
136.4 C 225kPa 6:43
136.0 C 221kPa 7:13
136.0 C 221kPa 7:13
VENTING 7:13
AIR DRYING 8:09
CYCLE COMPLETE 8:23

Digital Signature #
EBB3B997120B44C2
