
Midmark M11 - v1.0.5
Total Cycles: 3872

Name: RITTER2
Sterilizer ID: V2313228
08/19/2024 11:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 106.6 F 0.0
2:00 107.8 F 0.1
4:00 119.6 F 1.2
6:00 149.8 F 3.9
8:00 175.0 F 7.1
10:00 189.9 F 10.7
12:00 208.2 F 6.4
14:00 218.1 F 7.4
16:00 228.5 F 11.4
18:00 229.3 F 7.6
20:00 240.9 F 12.8
22:00 240.2 F 11.2
24:00 251.4 F 16.8
26:00 261.6 F 22.6
28:00 271.1 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.3 F 29.7
2:00 271.3 F 29.5
3:00 272.2 F 29.7
4:00 271.3 F 29.4

Min 271.1 F 29.2
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.4 F 0.0
5:00 186.0 F 0.0
10:00 182.8 F 0.0
15:00 176.0 F 0.0
20:00 172.2 F 0.0
25:00 170.4 F 0.0
30:00 170.1 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 28:44
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:07:20
CYCLE COMPLETE

08/19/2024 12:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____