
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000536
10:45 25/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:45
PRESSURIZING 3:48
STERILIZING 4:58
136.0 C 228kPa 4:58
136.2 C 229kPa 5:28
136.2 C 231kPa 5:58
136.3 C 232kPa 6:28
136.4 C 233kPa 6:58
136.2 C 230kPa 7:28
136.4 C 233kPa 7:58
136.4 C 233kPa 8:28
136.4 C 233kPa 8:28
VENTING 8:28
AIR DRYING 9:21
CYCLE COMPLETE 24:22

Digital Signature #
999C0BA82A659AF9
