

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5104

Name: RITTER 1  
Sterilizer ID: V2178198  
11/18/2024 03:56 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 97.2 F 0.0  
2:00 103.7 F 0.0  
4:00 131.0 F 0.1  
6:00 160.5 F 0.4  
8:00 188.1 F 0.7  
10:00 208.2 F 3.3  
12:00 220.4 F 6.5  
14:00 229.3 F 10.0  
16:00 234.5 F 10.1  
18:00 237.9 F 10.3  
20:00 236.0 F 8.6  
22:00 246.6 F 13.5  
24:00 244.6 F 12.3  
26:00 256.6 F 18.6  
28:00 267.2 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.0  
1:00 272.6 F 28.8  
2:00 271.7 F 28.1  
3:00 271.8 F 28.4  
4:00 272.6 F 28.7

Min 271.4 F 27.9  
Max 272.6 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.7 F 0.0  
5:00 216.1 F 0.0  
10:00 206.3 F 0.0  
15:00 201.1 F 0.0  
20:00 198.2 F 0.0  
25:00 195.2 F 0.0  
30:00 197.9 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 29:32  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:07:59  
CYCLE COMPLETE

11/18/2024 05:03 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_