
Midmark M11 - v1.0.0
Total Cycles: 763

Name: RITTER 1
Sterilizer ID: V2178198
03/31/2021 02:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.7 F 0.0
2:00 175.0 F 3.4
4:00 205.5 F 9.6
6:00 220.9 F 5.8
8:00 238.0 F 13.0
10:00 239.8 F 11.2
12:00 239.6 F 10.3
14:00 254.9 F 18.1
16:00 267.9 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 271.9 F 28.5
2:00 272.4 F 28.9
3:00 272.6 F 29.3
4:00 271.4 F 28.5

Min 271.4 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.8 F 0.0
5:00 238.4 F 0.0
10:00 243.5 F 0.0
Dry Cycle Aborted

FILLING: 2:06
HEATING: 17:07
STERILIZING: 4:00
VENTING: 2:55
DRYING: 10:13
TOTAL CYCLE: 00:36:21

03/31/2021 02:59 PM
ACCEPTED: X
REJECTED:
Operator: U01