
Midmark M11 - v1.0.5
Total Cycles: 1523

Name: RITTER2
Sterilizer ID: V2313228
09/22/2022 10:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.6 F 0.0
2:00 198.2 F 3.3
4:00 204.2 F 8.6
6:00 219.4 F 9.9
8:00 231.1 F 10.2
10:00 236.7 F 8.5
12:00 243.2 F 12.0
14:00 244.6 F 13.3
16:00 256.9 F 19.9
18:00 267.8 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 272.8 F 29.9
2:00 271.8 F 29.1
3:00 272.9 F 30.1
4:00 272.3 F 29.5

Min 271.7 F 29.1
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.4 F 0.0
5:00 209.1 F 0.0
10:00 218.0 F 0.0
15:00 219.4 F 0.0
20:00 219.4 F 0.0
25:00 219.2 F 0.0
30:00 219.0 F 0.0
Drying Complete

FILLING: 3:21
HEATING: 19:22
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:56
CYCLE COMPLETE

09/22/2022 11:12 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____