
Midmark M11 - v1.0.5
Total Cycles: 1595

Name: RITTER2
Sterilizer ID: V2313228
10/13/2022 12:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.5 F 0.0
2:00 184.2 F 3.5
4:00 192.7 F 8.6
6:00 214.3 F 9.2
8:00 228.7 F 9.6
10:00 232.7 F 8.9
12:00 242.7 F 13.3
14:00 242.9 F 12.5
16:00 255.9 F 19.3
18:00 267.4 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.4
1:00 272.7 F 29.8
2:00 271.9 F 29.5
3:00 272.8 F 29.8
4:00 271.9 F 29.3

Min 271.8 F 29.2
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.0 F 0.0
5:00 221.1 F 0.0
10:00 235.7 F 0.0
15:00 229.1 F 0.0
20:00 214.5 F 0.0
25:00 223.3 F 0.0
30:00 184.2 F 0.0
Drying Complete

FILLING: 3:01
HEATING: 19:23
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:36
CYCLE COMPLETE

10/13/2022 01:24 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____