

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3382

Name: RITTER2  
Sterilizer ID: V2313228  
02/13/2024 02:59 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 115.8 F 0.0  
2:00 117.2 F 0.3  
4:00 141.6 F 2.4  
6:00 169.0 F 6.1  
8:00 188.1 F 10.6  
10:00 211.2 F 4.8  
12:00 222.5 F 8.6  
14:00 235.2 F 13.5  
16:00 237.0 F 10.8  
18:00 235.7 F 9.2  
20:00 250.2 F 16.0  
22:00 262.4 F 23.0  
24:00 272.5 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.1  
1:00 272.2 F 29.4  
2:00 272.1 F 29.8  
3:00 271.8 F 29.2  
4:00 272.4 F 29.9

Min 271.3 F 29.0  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.4 F 0.0  
5:00 217.2 F 0.0  
10:00 201.6 F 0.0  
15:00 198.7 F 0.0  
20:00 196.5 F 0.0  
25:00 194.9 F 0.0  
30:00 192.9 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 24:14  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:02:30  
CYCLE COMPLETE

02/13/2024 04:01 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_