
Midmark M11 - v1.0.5
Total Cycles: 3346

Name: RITTER2
Sterilizer ID: V2313228
02/01/2024 02:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.3 F 0.0
2:00 122.0 F 0.7
4:00 141.7 F 3.3
6:00 177.8 F 7.5
8:00 198.4 F 12.6
10:00 220.5 F 6.6
12:00 233.1 F 11.6
14:00 233.0 F 8.6
16:00 239.8 F 10.6
18:00 246.1 F 13.7
20:00 258.8 F 20.6
22:00 270.4 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.0
1:00 272.1 F 29.4
2:00 271.5 F 29.3
3:00 272.3 F 29.3
4:00 271.3 F 29.3

Min 271.0 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.3 F 0.0
5:00 236.7 F 0.0
10:00 187.6 F 0.0
15:00 183.3 F 0.0
20:00 180.2 F 0.0
25:00 179.5 F 0.0
30:00 179.6 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 22:50
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:01:04
CYCLE COMPLETE

02/01/2024 03:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____