
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002095
7:47 01/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:45
STERILIZING 3:56
136.0 C 220kPa 3:56
136.2 C 222kPa 4:26
136.4 C 224kPa 4:56
136.4 C 224kPa 5:26
136.3 C 223kPa 5:56
136.1 C 222kPa 6:26
136.1 C 222kPa 6:56
136.1 C 220kPa 7:26
136.1 C 220kPa 7:26
VENTING 7:26
AIR DRYING 8:18
CYCLE COMPLETE 38:18

Digital Signature #
0202ABDD4AEDA9AD
