
Midmark M11 - v1.0.5
Total Cycles: 2382

Name: RITTER2
Sterilizer ID: V2313228
05/17/2023 08:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.6 F 0.0
2:00 175.5 F 1.0
4:00 180.9 F 5.0
6:00 200.6 F 10.0
8:00 222.5 F 6.4
10:00 228.8 F 10.4
12:00 232.0 F 7.8
14:00 242.1 F 11.9
16:00 245.4 F 13.4
18:00 258.1 F 20.3
20:00 269.5 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.6 F 29.7
2:00 271.9 F 29.0
3:00 271.7 F 29.4
4:00 272.5 F 29.7

Min 271.4 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.0 F 0.0
5:00 213.9 F 0.0
10:00 224.8 F 0.0
15:00 227.3 F 0.0
20:00 227.5 F 0.0
25:00 225.2 F 0.0
30:00 223.4 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 21:02
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:23
CYCLE COMPLETE

05/17/2023 09:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____