
Midmark M11 - v1.0.0
Total Cycles: 3892

Name: RITTER 1
Sterilizer ID: V2178198
10/12/2023 07:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.6 F 0.0
2:00 84.6 F 0.0
4:00 86.0 F 0.0
6:00 131.4 F 0.1
8:00 162.0 F 0.9
10:00 185.3 F 4.3
12:00 213.8 F 7.5
14:00 227.2 F 10.9
16:00 228.1 F 6.6
18:00 231.7 F 8.3
20:00 241.3 F 12.5
22:00 236.0 F 8.9
24:00 240.7 F 10.6
26:00 244.5 F 12.4
28:00 256.5 F 18.6
30:00 266.9 F 25.1
32:00 272.9 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 271.8 F 28.4
2:00 272.6 F 28.7
3:00 272.8 F 29.0
4:00 272.0 F 29.1

Min 271.2 F 28.4
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.1 F 0.0
5:00 235.4 F 0.0
10:00 223.7 F 0.0
15:00 221.2 F 0.0
20:00 180.5 F 0.0
25:00 178.5 F 0.0
30:00 177.2 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 32:13
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:10:22
CYCLE COMPLETE

10/12/2023 09:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____