
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001599
10:41 16/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:39
PRESSURIZING 2:43
STERILIZING 3:38
136.0 C 220kPa 3:38
136.4 C 225kPa 4:08
136.3 C 224kPa 4:38
136.4 C 226kPa 5:08
136.4 C 224kPa 5:38
136.4 C 223kPa 6:08
136.4 C 225kPa 6:38
136.4 C 222kPa 7:08
136.4 C 222kPa 7:08
VENTING 7:08
AIR DRYING 8:00
CYCLE COMPLETE 8:15

Digital Signature #
B4C42DAF96426E57
