
Midmark M11 - v1.0.0
Total Cycles: 3466

Name: RITTER 1
Sterilizer ID: V2178198
06/29/2023 11:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.7 F 0.0
2:00 195.6 F 3.6
4:00 207.6 F 8.2
6:00 225.3 F 10.0
8:00 231.2 F 9.6
10:00 238.8 F 9.2
12:00 244.5 F 13.1
14:00 243.7 F 12.1
16:00 256.5 F 18.7
18:00 267.5 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 29.0
2:00 271.8 F 28.3
3:00 271.7 F 28.6
4:00 272.6 F 29.0

Min 271.4 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.1 F 0.0
5:00 229.4 F 0.0
10:00 236.9 F 0.0
15:00 235.1 F 0.0
20:00 226.0 F 0.0
25:00 224.2 F 0.0
30:00 224.3 F 0.0
Drying Complete

FILLING: 2:03
HEATING: 19:26
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:47
CYCLE COMPLETE

06/29/2023 12:54 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____