
Midmark M11 - v1.0.0
Total Cycles: 5459

Name: RITTER 1
Sterilizer ID: V2178198
03/28/2025 08:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.1 F 0.0
2:00 76.6 F 0.4
4:00 88.0 F 1.9
6:00 127.4 F 4.5
8:00 155.6 F 7.3
10:00 174.4 F 10.6
12:00 185.5 F 4.1
14:00 205.4 F 4.1
16:00 216.0 F 7.4
18:00 228.1 F 11.2
20:00 227.7 F 6.1
22:00 239.9 F 11.3
24:00 236.4 F 8.8
26:00 249.5 F 15.0
28:00 260.5 F 21.1
30:00 270.9 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.8 F 28.9
2:00 271.7 F 28.1
3:00 272.9 F 29.0
4:00 272.2 F 28.4

Min 271.5 F 28.1
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.9 F 0.0
5:00 182.0 F 0.0
10:00 186.6 F 0.0
15:00 191.6 F 0.0
20:00 196.5 F 0.0
25:00 198.7 F 0.0
30:00 198.6 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 30:46
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:09:16
CYCLE COMPLETE

03/28/2025 09:35 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____