
Midmark M11 - v1.0.5
Total Cycles: 427

Name: RITTER2
Sterilizer ID: V2313228
10/07/2021 01:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.8 F 0.0
2:00 169.2 F 3.0
4:00 176.4 F 7.3
6:00 200.8 F 12.7
8:00 221.9 F 7.0
10:00 236.0 F 12.2
12:00 235.7 F 9.6
14:00 238.4 F 9.5
16:00 248.5 F 14.8
18:00 260.4 F 21.5
20:00 271.2 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 28.8
1:00 272.6 F 29.4
2:00 272.2 F 29.3
3:00 272.2 F 28.9
4:00 272.8 F 29.6

Min 271.8 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.6 F 0.0
5:00 232.1 F 0.0
10:00 222.6 F 0.0
15:00 224.4 F 0.0
20:00 212.4 F 0.0
25:00 231.6 F 0.0
30:00 218.6 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 20:32
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:21
CYCLE COMPLETE

10/07/2021 02:32 PM
ACCEPTED: X
REJECTED:
Operator: U02