
Midmark M11 - v1.0.5
Total Cycles: 379

Name: RITTER2
Sterilizer ID: V2313228
09/21/2021 06:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.4 F 0.0
2:00 74.4 F 0.0
4:00 77.4 F 0.0
6:00 92.4 F 0.2
8:00 129.1 F 1.6
10:00 157.0 F 5.2
12:00 182.3 F 8.9
14:00 200.7 F 12.7
16:00 212.5 F 6.6
18:00 224.5 F 10.9
20:00 231.0 F 7.8
22:00 243.1 F 13.1
24:00 241.3 F 11.7
26:00 253.3 F 17.8
28:00 263.8 F 24.1
30:00 271.9 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.3
1:00 272.4 F 30.2
2:00 272.5 F 30.1
3:00 272.5 F 30.1
4:00 272.5 F 30.1

Min 271.0 F 29.3
Max 273.1 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.7 F 0.0
5:00 229.8 F 0.0
10:00 234.3 F 0.0
15:00 230.5 F 0.0
20:00 203.5 F 0.0
25:00 189.4 F 0.0
30:00 175.5 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 30:09
STERILIZING: 4:00
VENTING: 2:60
DRYING: 30:00
TOTAL CYCLE: 01:08:07
CYCLE COMPLETE

09/21/2021 07:42 AM
ACCEPTED: X
REJECTED:
Operator: U01