
Midmark M11 - v1.0.5
Total Cycles: 768

Name: RITTER2
Sterilizer ID: V2313228
02/08/2022 09:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 97.4 F 0.0
2:00 101.7 F 0.0
4:00 120.7 F 0.1
6:00 156.3 F 1.9
8:00 181.0 F 5.5
10:00 201.2 F 9.7
12:00 216.5 F 7.9
14:00 227.6 F 9.6
16:00 233.7 F 9.4
18:00 241.8 F 12.6
20:00 241.7 F 11.6
22:00 254.5 F 18.2
24:00 265.9 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.4 F 29.5
2:00 271.5 F 29.0
3:00 272.6 F 29.7
4:00 271.8 F 28.9

Min 271.4 F 28.9
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.5 F 0.0
5:00 211.8 F 0.0
10:00 219.5 F 0.0
15:00 228.6 F 0.0
20:00 226.3 F 0.0
25:00 224.7 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 25:39
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:03:41
CYCLE COMPLETE

02/08/2022 10:48 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____