
Midmark M11 - v1.0.0
Total Cycles: 2331

Name: RITTER 1
Sterilizer ID: V2178198
09/15/2022 11:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.9 F 0.0
2:00 165.5 F 3.1
4:00 177.2 F 6.9
6:00 201.9 F 12.1
8:00 224.0 F 6.7
10:00 235.7 F 11.9
12:00 234.7 F 8.7
14:00 240.7 F 10.1
16:00 246.1 F 13.3
18:00 258.5 F 20.0
20:00 269.7 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.9 F 29.2
2:00 272.6 F 28.7
3:00 271.2 F 28.2
4:00 271.9 F 28.8

Min 271.1 F 28.1
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.9 F 0.0
5:00 232.1 F 0.0
10:00 212.7 F 0.0
15:00 205.3 F 0.0
20:00 203.3 F 0.0
25:00 215.8 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 21:02
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:38
CYCLE COMPLETE

09/15/2022 12:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____