

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2670

Name: RITTER2  
Sterilizer ID: V2313228  
07/19/2023 02:04 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.8 F 0.0  
2:00 177.3 F 3.1  
4:00 189.8 F 7.9  
6:00 212.4 F 13.3  
8:00 222.7 F 8.6  
10:00 234.0 F 10.9  
12:00 239.7 F 11.9  
14:00 239.8 F 10.8  
16:00 253.3 F 17.6  
18:00 265.0 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.7 F 29.9  
2:00 272.1 F 29.2  
3:00 271.6 F 29.4  
4:00 272.6 F 29.9  
  
Min 271.4 F 28.9  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.1 F 0.0  
5:00 213.5 F 0.0  
10:00 222.7 F 0.0  
15:00 222.7 F 0.0  
20:00 220.9 F 0.0  
25:00 222.3 F 0.0  
30:00 229.4 F 0.0  
  
Drying Complete

FILLING: 1:48  
HEATING: 19:50  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:53  
CYCLE COMPLETE

07/19/2023 03:02 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_