
Midmark M11 - v1.0.5
Total Cycles: 486

Name: RITTER2
Sterilizer ID: V2313228
10/26/2021 02:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.0 F 0.0
2:00 144.3 F 1.9
4:00 152.3 F 5.3
6:00 177.2 F 9.9
8:00 210.6 F 6.6
10:00 221.1 F 10.5
12:00 227.8 F 7.3
14:00 243.9 F 13.5
16:00 243.2 F 12.1
18:00 255.4 F 18.5
20:00 266.1 F 25.3
22:00 273.0 F 30.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.9
1:00 272.3 F 30.0
2:00 271.6 F 30.2
3:00 271.6 F 30.0
4:00 271.8 F 29.9

Min 271.2 F 29.8
Max 272.6 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 236.9 F 0.0
10:00 222.8 F 0.0
15:00 225.6 F 0.0
20:00 225.1 F 0.0
25:00 230.8 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 22:15
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:00:22
CYCLE COMPLETE

10/26/2021 03:40 PM
ACCEPTED: X
REJECTED:
Operator: U01