
Midmark M11 - v1.0.5
Total Cycles: 396

Name: RITTER2
Sterilizer ID: V2313228
09/27/2021 10:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 210.1 F 0.0
2:00 206.1 F 3.5
4:00 213.3 F 9.3
6:00 224.5 F 7.6
8:00 234.9 F 11.5
10:00 236.5 F 9.8
12:00 238.2 F 9.0
14:00 251.2 F 16.4
16:00 263.5 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.2 F 29.0
2:00 272.8 F 29.5
3:00 272.1 F 29.5
4:00 272.0 F 29.0

Min 271.5 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.9 F 0.0
5:00 224.7 F 0.0
10:00 222.9 F 0.0
15:00 223.3 F 0.0
20:00 222.5 F 0.0
25:00 223.2 F 0.0
30:00 228.7 F 0.0

Drying Complete

FILLING: 2:20
HEATING: 17:57
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:57:24
CYCLE COMPLETE

09/27/2021 11:25 AM
ACCEPTED: X
REJECTED:
Operator: U01