
Midmark M11 - v1.0.5
Total Cycles: 3196

Name: RITTER2
Sterilizer ID: V2313228
12/08/2023 10:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.7 F 0.0
2:00 72.4 F 0.0
4:00 81.0 F 0.2
6:00 94.3 F 1.1
8:00 131.0 F 3.5
10:00 162.4 F 6.2
12:00 183.9 F 9.2
14:00 195.4 F 13.2
16:00 211.6 F 5.6
18:00 221.0 F 9.2
20:00 231.0 F 13.2
22:00 233.0 F 9.3
24:00 239.0 F 10.6
26:00 243.4 F 12.7
28:00 254.6 F 18.5
30:00 264.9 F 24.7
32:00 271.1 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 272.5 F 29.9
2:00 271.6 F 29.2
3:00 271.8 F 29.7
4:00 272.5 F 29.7

Min 271.1 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.8 F 0.0
5:00 229.5 F 0.0
10:00 206.3 F 0.0
15:00 187.1 F 0.0
20:00 182.4 F 0.0
25:00 184.6 F 0.0
30:00 197.2 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 32:01
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:10:15
CYCLE COMPLETE

12/08/2023 11:18 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____