
Midmark M11 - v1.0.5
Total Cycles: 2586

Name: RITTER2
Sterilizer ID: V2313228
06/29/2023 10:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.3 F 0.0
2:00 184.7 F 3.4
4:00 194.7 F 8.2
6:00 216.3 F 12.3
8:00 227.4 F 9.3
10:00 236.1 F 9.3
12:00 242.6 F 13.1
14:00 242.7 F 12.2
16:00 255.9 F 19.1
18:00 267.5 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.7 F 29.9
2:00 272.0 F 29.2
3:00 271.5 F 29.3
4:00 272.6 F 29.9

Min 271.3 F 28.9
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 215.3 F 0.0
10:00 219.8 F 0.0
15:00 223.3 F 0.0
20:00 223.4 F 0.0
25:00 225.6 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 19:25
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:20
CYCLE COMPLETE

06/29/2023 11:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____