
Midmark M11 - v1.0.5
Total Cycles: 4492

Name: RITTER2
Sterilizer ID: V2313228
04/07/2025 11:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 105.7 F 0.0
2:00 103.5 F 0.5
4:00 106.4 F 2.3
6:00 132.2 F 5.1
8:00 157.9 F 8.3
10:00 183.7 F 12.0
12:00 213.3 F 3.9
14:00 212.1 F 7.5
16:00 224.1 F 11.4
18:00 228.9 F 6.9
20:00 240.0 F 12.0
22:00 238.4 F 10.1
24:00 250.1 F 15.8
26:00 260.5 F 21.7
28:00 270.0 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.1
1:00 272.5 F 30.1
2:00 271.6 F 29.9
3:00 271.5 F 29.5
4:00 272.3 F 29.8

Min 270.9 F 29.0
Max 273.3 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.4 F 0.0
5:00 240.4 F 0.0
10:00 216.9 F 0.0
15:00 209.4 F 0.0
20:00 165.5 F 0.0
25:00 159.1 F 0.0
30:00 157.4 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 28:50
STERILIZING: 4:00
VENTING: 4:01
DRYING: 30:00
TOTAL CYCLE: 01:07:44
CYCLE COMPLETE

04/07/2025 12:54 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____