
Midmark M11 - v1.0.0
Total Cycles: 2895

Name: RITTER 1
Sterilizer ID: V2178198
02/14/2023 08:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.0 F 0.0
2:00 80.0 F 0.0
4:00 87.1 F 0.4
6:00 120.5 F 2.5
8:00 148.1 F 5.2
10:00 175.4 F 8.5
12:00 188.8 F 12.2
14:00 207.3 F 3.6
16:00 217.6 F 7.2
18:00 227.8 F 11.0
20:00 229.2 F 7.0
22:00 240.9 F 11.7
24:00 237.0 F 9.2
26:00 249.6 F 15.2
28:00 260.4 F 21.2
30:00 270.5 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 271.8 F 28.9
2:00 272.0 F 28.7
3:00 271.3 F 28.5
4:00 272.2 F 29.1

Min 271.2 F 28.2
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.1 F 0.0
5:00 222.8 F 0.0
10:00 213.6 F 0.0
15:00 209.0 F 0.0
20:00 201.6 F 0.0
25:00 217.4 F 0.0
30:00 228.3 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 30:43
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:08:44
CYCLE COMPLETE

02/14/2023 09:57 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____