
Midmark M11 - v1.0.0
Total Cycles: 1008

Name: RITTER 1
Sterilizer ID: V2178198
07/14/2021 08:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.5 F 0.0
2:00 185.4 F 2.7
4:00 202.5 F 7.3
6:00 221.2 F 12.4
8:00 227.3 F 7.6
10:00 240.2 F 13.2
12:00 240.0 F 10.9
14:00 238.0 F 9.4
16:00 251.7 F 16.0
18:00 262.9 F 22.5
20:00 272.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.2
1:00 272.2 F 28.9
2:00 272.2 F 28.6
3:00 271.4 F 28.3
4:00 272.6 F 29.0

Min 271.4 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.8 F 0.0
5:00 229.8 F 0.0
10:00 215.8 F 0.0
15:00 209.8 F 0.0
20:00 196.0 F 0.0
25:00 203.7 F 0.0
30:00 214.8 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 20:14
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 00:59:26
CYCLE COMPLETE

07/14/2021 09:44 AM
ACCEPTED: X
REJECTED:
Operator: U01