
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000497
9:15 07/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:29
PRESSURIZING 3:31
STERILIZING 4:45
136.2 C 229kPa 4:45
136.2 C 230kPa 5:15
136.4 C 233kPa 5:45
136.4 C 229kPa 6:15
136.0 C 230kPa 6:45
136.1 C 229kPa 7:15
136.2 C 231kPa 7:45
136.2 C 232kPa 8:15
136.2 C 232kPa 8:15
VENTING 8:15
AIR DRYING 9:07
CASSETTE REMOVED
CYCLE COMPLETE 14:19

Digital Signature #
418A3A7B0EA93841
