
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001537
12:42 10/28/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:17
PRESSURIZING 7:13
STERILIZING 9:00
136.0 C 220kPa 9:00
135.7 C 217kPa 9:30
136.3 C 219kPa 10:00
136.3 C 221kPa 10:30
136.4 C 220kPa 11:00
136.3 C 221kPa 11:30
136.0 C 220kPa 12:00
136.4 C 219kPa 12:30
VENTING 12:30
AIR DRYING 13:35
CYCLE COMPLETE 34:15

Digital Signature #
2014D9678D4E9133
