

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1240

Name: RITTER 1  
Sterilizer ID: V2178198  
10/07/2021 10:16 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.1 F 0.0  
2:00 165.1 F 0.8  
4:00 169.9 F 4.3  
6:00 194.9 F 8.8  
8:00 214.9 F 9.2  
10:00 227.1 F 7.2  
12:00 238.4 F 12.3  
14:00 237.4 F 9.2  
16:00 238.0 F 9.1  
18:00 248.4 F 14.2  
20:00 260.7 F 21.1  
22:00 272.2 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 27.9  
1:00 272.6 F 28.8  
2:00 272.0 F 28.2  
3:00 271.1 F 28.0  
4:00 272.7 F 28.8

Min 271.1 F 27.7  
Max 272.9 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.8 F 0.0  
5:00 234.2 F 0.0  
10:00 218.1 F 0.0  
15:00 228.4 F 0.0  
20:00 226.8 F 0.0  
25:00 225.0 F 0.0  
30:00 211.5 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 22:33  
STERILIZING: 4:00  
VENTING: 2:58  
DRYING: 30:00  
TOTAL CYCLE: 01:00:37  
CYCLE COMPLETE

10/07/2021 11:16 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_