
Midmark M11 - v1.0.0
Total Cycles: 1216

Name: RITTER 1
Sterilizer ID: V2178198
09/29/2021 08:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.2 F 0.0
2:00 142.7 F 0.7
4:00 169.7 F 2.5
6:00 193.4 F 5.6
8:00 209.3 F 9.8
10:00 223.9 F 7.9
12:00 228.0 F 7.7
14:00 239.8 F 12.2
16:00 235.7 F 8.9
18:00 242.5 F 10.9
20:00 245.7 F 12.9
22:00 257.3 F 19.1
24:00 267.9 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.6 F 28.9
2:00 272.0 F 28.3
3:00 271.2 F 28.2
4:00 272.5 F 28.9

Min 271.2 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 233.7 F 0.0
10:00 226.6 F 0.0
15:00 222.4 F 0.0
20:00 216.3 F 0.0
25:00 214.5 F 0.0
Dry Cycle Aborted

FILLING: 0:59
HEATING: 25:26
STERILIZING: 4:00
VENTING: 3:19
DRYING: 29:39
TOTAL CYCLE: 01:03:23

09/29/2021 09:11 AM
ACCEPTED: X
REJECTED:
Operator: U01