
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000185
11:15 17/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:06
PRESSURIZING 4:06
STERILIZING 5:34
136.4 C 225kPa 5:34
136.4 C 227kPa 6:04
136.3 C 225kPa 6:34
136.4 C 226kPa 7:04
136.4 C 226kPa 7:34
136.4 C 224kPa 8:04
136.2 C 224kPa 8:34
136.3 C 225kPa 9:04
136.3 C 225kPa 9:04
VENTING 9:04
AIR DRYING 10:09
CYCLE COMPLETE 25:09

Digital Signature #
47262CC6474DE6B5
