
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 001094
14:37 25/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:45
STERILIZING 3:54
136.4 C 220kPa 3:54
136.2 C 224kPa 4:24
136.3 C 225kPa 4:54
136.1 C 222kPa 5:24
136.1 C 224kPa 5:54
136.1 C 224kPa 6:24
136.4 C 226kPa 6:54
136.4 C 226kPa 7:24
136.4 C 226kPa 7:24
VENTING 7:24
AIR DRYING 8:16
CYCLE COMPLETE 23:16

Digital Signature #
E890D05E867C5942
