
Midmark M11 - v1.0.5
Total Cycles: 1270

Name: RITTER2
Sterilizer ID: V2313228
07/07/2022 04:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.1 F 0.0
2:00 122.3 F 1.1
4:00 132.3 F 3.8
6:00 158.0 F 7.7
8:00 184.6 F 12.2
10:00 208.6 F 5.1
12:00 223.0 F 9.4
14:00 233.7 F 10.9
16:00 236.8 F 10.8
18:00 235.1 F 8.7
20:00 248.1 F 15.1
22:00 259.6 F 21.5
24:00 270.1 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.3
1:00 272.3 F 29.9
2:00 272.4 F 29.5
3:00 271.9 F 29.6
4:00 272.7 F 29.9

Min 271.6 F 29.1
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.6 F 0.0
5:00 214.3 F 0.0
10:00 227.4 F 0.0
15:00 218.6 F 0.0
20:00 218.7 F 0.0
25:00 217.5 F 0.0
30:00 218.1 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 24:46
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:03:23
CYCLE COMPLETE

07/07/2022 05:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____