

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 007231  
8:35 19/04/2024

WRAPPED

135 C for 10 min.

CYCLE START	0:00
CONDITIONING	2:26
PRESSURIZING	4:24
STERILIZING	5:23
136.5 C 222kPa	5:23
136.4 C 222kPa	5:53
136.2 C 222kPa	6:23
136.4 C 222kPa	6:53
136.4 C 221kPa	7:23
136.2 C 220kPa	7:53
136.4 C 220kPa	8:23
135.8 C 218kPa	8:53
136.2 C 221kPa	9:23
136.1 C 218kPa	9:53
136.1 C 219kPa	10:23
136.2 C 222kPa	10:53
136.1 C 220kPa	11:23
136.4 C 221kPa	11:53
136.2 C 219kPa	12:23
136.3 C 222kPa	12:53
136.4 C 221kPa	13:23
136.2 C 219kPa	13:53
136.3 C 222kPa	14:23
136.4 C 222kPa	14:53
136.4 C 221kPa	15:23
136.4 C 221kPa	15:23
VENTING	15:23
AIR DRYING	16:10
CYCLE COMPLETE	46:10

-----

Digital Signature #

4E9FD9879BD004C7

-----