

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 007704  
10:26 07/08/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:29  
PRESSURIZING 2:34  
STERILIZING 3:30  
136.0 C 220kPa 3:30  
136.4 C 222kPa 4:00  
136.4 C 222kPa 4:30  
136.4 C 224kPa 5:00  
136.4 C 223kPa 5:30  
136.4 C 222kPa 6:00  
136.1 C 217kPa 6:30  
136.4 C 224kPa 7:00  
136.4 C 224kPa 7:00  
VENTING 7:00  
AIR DRYING 7:56  
CYCLE COMPLETE 35:19  
-----

Digital Signature #  
D0B1E66ED85C85FD  
-----