

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 006235  
11:47 05/09/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:59  
PRESSURIZING 2:03  
STERILIZING 3:01  
136.0 C 220kPa 3:01  
136.4 C 225kPa 3:31  
136.4 C 222kPa 4:01  
136.4 C 225kPa 4:31  
136.4 C 225kPa 5:01  
136.4 C 224kPa 5:31  
136.4 C 225kPa 6:01  
136.3 C 224kPa 6:31  
136.3 C 224kPa 6:31  
VENTING 6:31  
AIR DRYING 7:21  
CYCLE COMPLETE 7:23

-----  
Digital Signature #  
3A84E5CD44017B13  
-----