
Midmark M11 - v1.0.0
Total Cycles: 4728

Name: RITTER 1
Sterilizer ID: V2178198
07/15/2024 11:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.6 F 0.0
2:00 168.2 F 2.9
4:00 170.1 F 6.3
6:00 191.1 F 10.2
8:00 216.4 F 6.9
10:00 218.2 F 7.4
12:00 230.6 F 11.7
14:00 231.0 F 7.7
16:00 242.9 F 13.0
18:00 241.4 F 11.3
20:00 253.2 F 17.2
22:00 263.5 F 23.2
24:00 272.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 271.6 F 28.9
2:00 272.7 F 29.1
3:00 271.6 F 28.6
4:00 271.2 F 28.7

Min 271.1 F 28.4
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.5 F 0.0
5:00 200.1 F 0.0
10:00 212.7 F 0.0
15:00 202.9 F 0.0
20:00 222.3 F 0.0
25:00 222.8 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 24:11
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:03:08
CYCLE COMPLETE

07/15/2024 12:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____