
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001781
10:02 22/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:56
PRESSURIZING 3:01
STERILIZING 4:13
136.0 C 220kPa 4:13
136.2 C 223kPa 4:43
136.4 C 221kPa 5:13
136.3 C 223kPa 5:43
136.4 C 224kPa 6:13
136.4 C 225kPa 6:43
136.2 C 222kPa 7:13
136.4 C 221kPa 7:43
136.4 C 221kPa 7:43
VENTING 7:43
AIR DRYING 8:33
CYCLE COMPLETE 11:45

Digital Signature #
AC23CA36CE7C2E3E
