
Midmark M11 - v1.0.0
Total Cycles: 3156

Name: RITTER 1
Sterilizer ID: V2178198
04/24/2023 04:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.0 F 0.0
2:00 180.6 F 0.0
4:00 205.1 F 1.8
6:00 224.1 F 6.9
8:00 235.1 F 12.1
10:00 235.3 F 9.2
12:00 239.9 F 9.5
14:00 243.8 F 11.7
16:00 247.9 F 13.9
18:00 261.2 F 21.3
20:00 272.8 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.3 F 28.4
2:00 272.0 F 28.6
3:00 272.0 F 28.3
4:00 272.1 F 28.7

Min 271.5 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.0 F 0.0
5:00 225.8 F 0.0
10:00 225.4 F 0.0
15:00 230.5 F 0.0
20:00 224.0 F 0.0
25:00 221.5 F 0.0
30:00 220.0 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 20:20
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:35
CYCLE COMPLETE

04/24/2023 05:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____