

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1549

Name: RITTER2  
Sterilizer ID: V2313228  
10/03/2022 07:09 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.7 F 0.0  
2:00 79.4 F 0.0  
4:00 103.1 F 0.0  
6:00 109.3 F 0.0  
8:00 136.2 F 0.0  
10:00 171.3 F 0.6  
12:00 194.4 F 0.9  
14:00 210.5 F 2.9  
16:00 228.0 F 6.4  
18:00 234.8 F 10.2  
20:00 237.4 F 9.4  
22:00 243.2 F 12.4  
24:00 241.2 F 11.0  
26:00 239.4 F 10.1  
28:00 253.3 F 16.9  
30:00 264.8 F 23.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.4  
1:00 272.1 F 28.5  
2:00 272.6 F 29.0  
3:00 272.1 F 28.9  
4:00 271.8 F 28.3

Min 271.5 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 216.8 F 0.0  
5:00 216.7 F 0.0  
10:00 223.3 F 0.0  
15:00 222.9 F 0.0  
20:00 223.8 F 0.0  
25:00 222.1 F 0.0  
30:00 220.3 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 31:50  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:09:47  
CYCLE COMPLETE

10/03/2022 08:18 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_