

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1036

Name: RITTER2  
Sterilizer ID: V2313228  
04/26/2022 01:25 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.6 F 0.0  
2:00 169.5 F 2.6  
4:00 178.6 F 7.0  
6:00 212.2 F 12.4  
8:00 221.6 F 7.6  
10:00 237.6 F 13.0  
12:00 236.5 F 10.6  
14:00 236.6 F 9.3  
16:00 250.5 F 16.3  
18:00 262.4 F 23.3  
20:00 272.2 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.2  
1:00 271.9 F 29.9  
2:00 272.4 F 29.6  
3:00 271.1 F 29.2  
4:00 272.2 F 30.1

Min 271.1 F 29.1  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.9 F 0.0  
5:00 223.6 F 0.0  
10:00 221.0 F 0.0  
15:00 210.0 F 0.0  
20:00 191.7 F 0.0  
25:00 180.5 F 0.0  
30:00 176.7 F 0.0  
Drying Complete

FILLING: 1:46  
HEATING: 20:04  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:01  
CYCLE COMPLETE

04/26/2022 02:38 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02