
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000491
14:56 09/04/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:14
STERILIZING 4:32
136.3 C 231kPa 4:32
136.4 C 233kPa 5:02
136.2 C 231kPa 5:32
136.4 C 233kPa 6:02
136.4 C 233kPa 6:32
136.2 C 230kPa 7:02
136.4 C 232kPa 7:32
136.3 C 232kPa 8:02
136.3 C 232kPa 8:02
VENTING 8:02
AIR DRYING 8:57
CYCLE COMPLETE 9:17

Digital Signature #
501322CABF65136F
