
Midmark M11 - v1.0.0
Total Cycles: 2085

Name: RITTER 1
Sterilizer ID: V2178198
07/04/2022 04:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.1 F 0.0
2:00 165.5 F 3.4
4:00 187.8 F 8.0
6:00 206.7 F 9.3
8:00 225.7 F 7.8
10:00 238.7 F 13.1
12:00 239.4 F 10.5
14:00 236.6 F 8.7
16:00 252.1 F 16.0
18:00 264.7 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 27.9
1:00 272.6 F 28.6
2:00 271.5 F 28.0
3:00 272.3 F 28.3
4:00 271.8 F 28.3

Min 271.3 F 27.7
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.5 F 0.0
5:00 215.2 F 0.0
10:00 218.0 F 0.0
15:00 218.5 F 0.0
20:00 211.7 F 0.0
25:00 206.6 F 0.0
30:00 207.8 F 0.0

Drying Complete

FILLING: 1:04
HEATING: 19:37
STERILIZING: 4:00
VENTING: 2:56
DRYING: 30:00
TOTAL CYCLE: 00:57:38
CYCLE COMPLETE

07/04/2022 05:24 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____