
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007643
8:08 10/07/2024

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:39
STERILIZING 3:44
136.0 C 220kPa 3:44
135.8 C 219kPa 4:14
136.4 C 220kPa 4:44
136.4 C 222kPa 5:14
136.4 C 223kPa 5:44
136.1 C 221kPa 6:14
136.4 C 222kPa 6:44
136.4 C 223kPa 7:14
136.4 C 223kPa 7:14
VENTING 7:14
AIR DRYING 8:04
CYCLE COMPLETE 14:52

Digital Signature #
F58C4786026D7842
