
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003375
10:08 04/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:37
PRESSURIZING 3:38
STERILIZING 4:47
136.4 C 220kPa 4:47
136.2 C 221kPa 5:17
136.4 C 224kPa 5:47
136.4 C 223kPa 6:17
136.2 C 223kPa 6:47
136.3 C 222kPa 7:17
136.2 C 223kPa 7:47
136.4 C 220kPa 8:17
136.4 C 220kPa 8:17
VENTING 8:17
AIR DRYING 9:13
CYCLE COMPLETE 9:16

Digital Signature #
A2FDB33FDA8A0E23
