
Midmark M11 - v1.0.5
Total Cycles: 1223

Name: RITTER2
Sterilizer ID: V2313228
06/24/2022 03:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.3 F 0.0
2:00 167.7 F 3.1
4:00 184.2 F 7.1
6:00 205.7 F 12.1
8:00 226.6 F 7.4
10:00 233.5 F 12.8
12:00 233.7 F 10.0
14:00 235.0 F 8.7
16:00 247.4 F 14.9
18:00 259.3 F 21.4
20:00 269.9 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 271.9 F 30.0
2:00 272.7 F 30.2
3:00 272.5 F 29.9
4:00 271.9 F 29.6

Min 271.1 F 29.2
Max 272.9 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.3 F 0.0
5:00 212.6 F 0.0
10:00 221.3 F 0.0
15:00 226.1 F 0.0
20:00 229.4 F 0.0
25:00 226.7 F 0.0
30:00 183.9 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 21:03
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:57
CYCLE COMPLETE

06/24/2022 04:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____